



OPEN

Conical shell X-ray beam tomosynthesis and micro-computed tomography for microarchitectural characterisation

Emily L. Arnold^{1✉}, Farid Elarnaut², David Downes², J. Paul O. Evans², Charlene Greenwood³ & Keith D. Rogers¹

Bone quality is commonly used to diagnose bone diseases such as osteoporosis, with many studies focusing on microarchitecture for fracture prediction. In this study a bovine distal femur was imaged using both micro-computed tomography (μ CT) and tomosynthesis using focal construct geometry (FCG) for comparison of microarchitectural parameters. Six regions of interest (ROIs) were compared between the two imaging modalities, with both global and adaptive methods used to binarize the images. FCG images were downsampled to the same pixel size as the μ CT images. Bone morphometrics were determined using BoneJ, for each imaging modality, binarization technique and ROI. Bone area/total area was found to have few significant differences between FCG and μ CT ($p < 0.05$ for two of six ROIs). Fractal Dimension had only one significant difference ($p < 0.05$ for one of six ROIs) between μ CT and downsampled FCG (where pixel size was equalized). Trabecular thickness and trabecular spacing were observed to follow trends as observed for the corresponding μ CT images, although many absolute values were significantly different ($p < 0.05$ for between one and six ROIs depending on image types used). This study demonstrates the utility of tomosynthesis for measurement of microarchitectural morphometrics.

Abbreviations

μ CT	Micro-computed tomography
FCG	Focal construct geometry
ROI	Region of interest
BA/TA	Bone area/total area
FD	Fractal dimension
TbTh	Trabecular thickness
TbSp	Trabecular separation
DXA	Dual energy X-ray absorptiometry
BMD	Bone mineral density
HR-pQCT	High resolution peripheral quantitative CT
CBCT	Cone beam CT
MRI	Magnetic resonance imaging
XRD	X-ray diffraction
MSCT	Multislice CT

Osteoporosis is a sweeping condition which currently affects over 3.7 million people in the UK¹, and is estimated that the treatment and prevention will cost £5.5 billion annually by 2025². Currently, dual energy X-ray absorptiometry (DXA) is used as the 'gold standard' for diagnosis of osteoporosis³ by measuring bone mineral density

¹Cranfield Forensic Institute, Cranfield University, Shrivenham SN6 8LA, Wiltshire, UK. ²Imaging Science Group, Nottingham Trent University, Rosalind Franklin Building, Nottingham NG11 8NS, UK. ³School of Chemical and Physical Sciences, Keele University, Keele ST5 5BJ, Staffordshire, UK. ✉email: e.l.arnold@cranfield.ac.uk

(BMD) at the lumbar spine and hip. DXA is often used in conjunction with additional tools such as FRAX⁴ for improved fracture prediction. Bone strength is thought to be a better predictor for fracture than BMD alone, encompassing both microarchitectural and physicochemical 'bone quality'^{5–7}. In an effort to better quantify bone quality, through several different imaging techniques, a number of studies have shown that microarchitectural parameters change significantly in diseased bone compared to healthy bone, including trabecular thickness, trabecular separation, and fractal dimension^{8–12}.

Recently studies have investigated the use of high resolution peripheral quantitative computed tomography (HR-pQCT) for fracture prediction^{13,14}, cone beam CT (CBCT)¹⁵, and magnetic resonance imaging (MRI)¹⁶, often compared to micro-CT (μ CT). Currently, HR-pQCT can achieve the highest resolution, at an isotropic resolution of either 61 μ m or 82 μ m, while in practice it achieves a spatial resolution of 100 μ m or 142 μ m¹⁴. However, a considerable limitation of HR-pQCT remains that it can only be used to examine peripheral sites. Relatively recent studies have used MRI to image trabecular bone and achieved an in-plane resolution of 156 μ m with a slice thickness of 0.5 mm at the distal radius¹⁷, and an in-plane resolution of 234 μ m with a slice thickness of 1.5 mm at the hip¹⁶. In comparison, CBCT can achieve a voxel size between 75 μ m and 500 μ m¹⁵. Additionally, opportunistic screening using clinical CT is a relatively common topic within literature¹⁸, as is the use of more traditional digital tomosynthesis for fracture prediction¹⁹.

Focal construct geometry (FCG) is an alternative X-ray beam geometry which uses a conical shell X-ray beam, designed to reduce exposure time for collection. It also provides unique access to the coherent scatter signals, thus providing crystallographic detail of the inspected material. FCG has been applied to X-ray absorption imaging^{20,21} and X-ray diffraction (XRD) in energy dispersive²² and angular dispersive modes^{23–25}, notably on bone^{22,26}. Recently this approach has been expanded to benefit from sporadic collection to further reduce dose and scanning time²¹. Additionally, a high-resolution technique using interleaved sampling has been applied to decrease pixel size in the xy plane²⁷ by an order of magnitude.

To evaluate the application of absorption FCG imaging within a clinical setting, and further the analytical methodology introduced previously²⁷, this work compares FCG to μ CT imaging quantified through measurement of trabecular microarchitecture morphometrics in a distal bovine femur. To evaluate the accuracy of microarchitectural parameters measured from absorption FCG imaging, six regions of interest (ROIs) were selected, and four parameters were measured for each imaging modality. This analysis demonstrates the ability of FCG to measure microarchitectural properties, potentially useful for clinical diagnostics in the future.

Methods

Sample

A distal bovine femur was sourced from a local butcher (18–24 months), approximately 115 mm $\times$ 144 mm $\times$ 99 mm.

μ CT

The bone sample was scanned using a Nikon CT H225 (X-Tek Systems Ltd, Tring, Hertfordshire, UK) cone beam μ -CT scanner operated at 125 kV and 101 μ A, with a voxel size of 125 μ m. Noise correction and beam hardening corrections were applied.

FCG

An FCG raster scan was performed similarly to that described previously^{20,21,27}. A Hamamatsu L1261-07 micro-focus X-ray source was used with an accelerating voltage and current of 150 kV and 67 μ A, respectively. Conical projections were collected on a Siemens HIDEQ 33-4 ISX image intensifier with a CsI phosphor input screen ($\sim$ 160 mm diameter useful entrance field). The intensifier is optically coupled to a Videomed-XR2 GigE camera (1024 $\times$ 1024). The distance between the detector surface and the X-ray point source was fixed at 409 mm, with the bone sample and translation stage fixed 139 mm from the X-ray point source, as shown in Fig. 1.

The rectilinear raster scan was performed using 501 $\times$ 501 points with a 0.3 mm step size giving a 150 mm $\times$ 150 mm field of view for the digital tomosynthesis. A series of 18 concentric conical shell projections were measured at each scan step, with the mean annular ring radii linearly spaced from $r = 7.89$ mm to 75 mm, giving each a mean half opening from 1.11° to 10.39°, according to $\tan^{-1}(r/z)$ where $z = 409$ mm.

A digital tomogram or sectional image was reconstructed from the ring projections using sparse interleaved sampling with 4 $\times$ upscaling described in detail elsewhere²⁷. Applying this method produced an image stack with 75 μ m pixels in both x and y axes, with a slice thickness of 1 mm in z .

Registration

To enable comparison of μ CT and FCG images, a single slice of the FCG image stack was chosen for registration and three unique features were chosen within this slice (shown in Fig. 2). The μ CT volume was rotated about each of its three orthogonal axes until the features chosen in the FCG slice were all in focus in one plane of the μ CT volume. The uniqueness of the section was verified by examination of the entirety of both the FCG and μ CT volumes. The accuracy of the relative in-plane registration was confirmed by comparative one-to-one coordinate analysis. The relative axial separations between the features were all found to be in good agreement, and within the limits of the predicted spatial resolution. Thus, the chosen features acted as surrogate fiducial markers of a unique plane within the experimental limits. The sectional image arising from this plane was used for all subsequent analysis.

Dissimilar z -axes resolution or section thickness produces an increase in FCG sectional image depth containing relatively more in-plane trabeculae.




Figure 1. Instrument configuration. The bone sample sits on a translation stage, which enables a two axis x,y raster scan relative to the hollow beam. This arrangement is equivalent geometrically to raster scanning the beam over a stationary bone sample.




Figure 2. Landmarks used for registration of FCG (left) and μ CT (right) slices.

To enable like-for-like comparison of μ CT and FCG images, the FCG slice was downsampled to the same pixel size as the μ CT image ($125\ \mu\text{m}$) using the Matlab²⁸ function *imresize* with bicubic interpolation. All further processing and analysis will be carried out on both the original FCG and downsampled FCG images.

Feature exaggeration and artefact suppression

Blur arising from out-of-plane structures is a well-known difficulty when using shift-and-add reconstruction for tomosynthesis²⁹ and is readily apparent within the FCG images seen here. When thresholding is attempted (either adaptive or global) poor results are seen where many trabeculae were not present, while large areas of cartilage and blur from out-of-focus planes are represented in the foreground (Fig. S1 in Supplemental Information).

While there are several deblurring algorithms and alternative reconstruction algorithms used within clinical tomosynthesis applications previously²⁹, to combat these artefacts within this research a morphological top-hat filter was applied to reduce uneven background intensity. Top-hat (or white top-hat) transforms have been

used in the past for clinical imaging purposes, for example for whole body sintigraphy³⁰ and for examination of pathological calcifications^{31,32}. To ensure consistent processing, both FCG and μ CT images underwent the same processing procedure. All processing steps were carried out in Matlab²⁸ and are given below:

- A disk structuring element was created with either $r = 5$ pixels (625 μ m) for μ CT and downsampled FCG or $r = 9$ pixels (675 μ m) for FCG
- *imtophat* was used to morphologically open the image (erosion followed by dilation using the specified structuring element) then subtract the opened image from the original image
- *imadjust* was used to adjust image intensity values, saturating the upper 1% and lower 1% of the image.

The structuring element size was chosen to be larger than the features being targeted (trabecular thickness and separation). In this case, trabecular thickness in bovine trabecular bone is commonly reported between 110 and 600 μ m, while trabecular separation is between 60 and 600 μ m^{33–37}. It should be noted though that most studies observe the femoral head or proximal tibia rather than the distal femur, and TbTh and TbSp have been seen to vary significantly based on the skeletal site and model used for calculation³³.

The initial FCG image and processed FCG image are shown in Fig. 3a, b, respectively, while the initial μ CT image and processed μ CT image are shown in Fig. 3c, d, respectively. Initial images were not used for any further analysis, and only processed images will be used from this point forward. Within the images, it should be noted that most cortical bone ‘fades’ with the top-hat transform, as it is significantly larger than the structuring element used. This illustrates the importance in the choice of structuring element, particularly the radius (size).

Binarization

Two binarization methods were used, adaptive and global, for each of the three images (μ CT, FCG and downsampled FCG). Adaptive thresholding was applied using Bradley’s method³⁸ with a neighbourhood size of either 151 $\times$ 151 pixels (μ CT and downsampled FCG) or 251 $\times$ 251 pixels (FCG); the sensitivity value s is user specified based on visual inspection. For global thresholding, the threshold value t is specified by the user based on visual inspection. Figure 4 shows both adaptive and global binarizations for μ CT and FCG images. An inset image shows a section of particular interest (due to the clearly visible landmarks) in greater detail. An apparent artifact can be seen in Fig. 4a, due to the top-hat transformation operating on a noisy background.

Morphometric measurement

To interrogate accuracy of microarchitectural morphometrics, six regions of interest (ROIs) were selected in areas where trabecular bone was clearly visible for both global and adaptive binarizations of μ CT and FCG images. Each region of interest was set as either 50 $\times$ 50 pixels (for μ CT & downsampled FCG) or 83 $\times$ 83 (for FCG), seen in Fig. 5a, b for FCG and μ CT, respectively.

To test repeatability, each ROI was shifted to eight positions around the selected ROI ($\pm 20\%$ of ROI size), to produce a total of nine images which contributed to the morphometric measurement for each ROI (illustrated in Fig. 6). Mean and standard deviation was calculated for each set of nine images.




Figure 3. (a) initial and (b) processed FCG image, (c) initial and (d) processed μ CT image.




Figure 4. Binarized images for (a) adaptive thresholding of the FCG image, (b) global thresholding of the FCG image, (c) adaptive thresholding of the μ CT image and (d) global thresholding of the μ CT image.




Figure 5. Selected regions of interest in the global binarization of (a) the FCG image and (b) the μ CT image.

Statistical analysis

Several morphometric parameters were determined using BoneJ^{39,40} in Fiji⁴¹: bone area/total area (BA/TA), fractal dimension (FD), trabecular thickness (TbTh) and trabecular separation (TbSp). Calculation of both TbTh and TbSp is not dependent upon either a plate or rod model but fits maximal circles to the structures in a model-independent method⁴². For each ROI in each of the six image types (both global and adaptive binarizations of μ CT, FCG and downsampled FCG), each parameter measured was tested for normality using either a Shapiro–Wilk or a Shapiro–Francia test, depending on the kurtosis of the sample⁴³. To determine significant differences between image type, if both samples being compared are normal, a two-sample T-test is performed assuming unequal variances; if one or both samples being compared are non-normal a Wilcoxon rank sum test was performed. In both cases, significance was set at $p = 0.05$.

Results

Qualitative comparison

Images from each ROI were compared for each imaging modality and each binarization method, shown in Fig. 7. While the difference in z resolution between the FCG image stack and μ CT volume complicates the precise registration of individual trabeculae, consistent patterns within the trabecular structure can be observed. Implications of adaptive and global thresholding are considered further in the discussion.




Figure 6. Example of the position of the nine sections taken from the adaptive binarization of the FCG image used which were analysed to produce data for region of interest one.




Figure 7. Regions of interest for each binarization method of each imaging technique.

Quantitative comparison

Microarchitectural morphometrics were compared for both global and adaptive binarizations for each imaging modality. Results for all morphometric characteristics are given in Fig. 8, and results of all significance tests are given in Fig. S2 in Supplemental Information.

BA/TA for each ROI is shown in Fig. 8a. BA/TA values ranged from ~0.40 to ~0.65, varying with location, imaging method and binarization method. Several outliers are immediately noted, most often when the global binarization was used. Most ROIs give few significant differences between global and adaptive binarization of μ CT images and taking the adaptive binarization of μ CT as ground truth it is apparent most FCG and downsampled FCG images follow general trends. Perhaps most notable is the lack of significant differences between the adaptive binarization of FCG and μ CT ($p < 0.05$ for only two of the six ROIs); this demonstrates the ability of FCG to measure BA/TA accurately compared to 'ground truth' μ CT.

Fractal dimension (FD) shows much larger discrepancies between images with different pixel sizes (Fig. 8b). FD for FCG images is significantly different from all μ CT and downsampled FCG images. In contrast, very few significant differences are seen between downsampled FCG and μ CT images, particularly when an adaptive binarization is used ($p < 0.05$ for only ROI 4 when adaptive downsampled FCG is compared to both adaptive and global μ CT). It is also of note that few significant differences are present when any adaptive binarization is compared to the global binarization from the same imaging mode (e.g. μ CT, FCG or downsampled FCG).

When considering mean trabecular thickness (Fig. 8c), general trends can be observed for most ROIs. Similarly, mean trabecular separation (Fig. 8d) shows general agreement, particularly for adaptive binarizations. Both TbTh and TbSp show less of a pattern when significant differences between image types are examined. Of note is the different binarization techniques for each imaging modality, where at least two to four ROIs are significantly




Figure 8. Morphometric parameters: (a) BA/TA, (b) fractal dimension, (c) mean TbTh and (d) mean TbSp. Error bars represent standard deviation for nine shifted images within each ROI.

different ($p < 0.05$) for each. This result indicates that binarization method is very important for TbTh and TbSp measurement, particularly for FCG.

Discussion

Thresholding has been seen to affect morphometric measurements in previous studies^{44,45}, and is also apparent here. Differences resulting from adaptive or global thresholding were apparent in some parameters (BA/TA, TbTh and TbSp) while they did not affect others (such as FD). Also, some image types and binarizations are clear outliers (for example, global binarizations of both original FCG and downsampled FCG for BA/TA of ROI 2 and ROI 5), and often explained by the visual inspection of the ROI; for global binarizations of FCG at ROI 2 a large void is present near the lower part of the region, most likely due to the drastic change in background intensity in the original image, resulting in a relatively rapid intensity change of the trabeculae images after processing. This can also be seen in Fig. 4a, b insets, which show the area around ROI 5. Similarly a void can be seen for both global and adaptive binarizations at ROI 3, potentially explaining the differences for ROI 3 across imaging modalities.

This disagreement is perhaps demonstrated better in ROI 2, while the trabeculae near the growth plate are visually apparent in processed FCG images, the difference in intensity clearly introduces issues for global binarization. Close inspection of areas such as ROI 2 and ROI 5 with a high background intensity (and thus a potentially lower intensity of trabeculae after the top-hat transform) shows that while adaptive binarization technique accommodates this change in intensity better than the global binarization technique, it still does not ideally threshold. There is potential if the size of the neighbourhood used in the adaptive binarization is reduced, that this may allow more accurate thresholding of trabeculae network images with large changes in intensity, but it would potentially include more noise elsewhere in the image. This can be seen already for adaptive binarizations of both μ CT and FCG images in Fig. 4a, c, where noise within the shaft of the femur has been discerned by the adaptive threshold due to the lack of intense features in the neighbourhood.

It is readily apparent that both global and adaptive binarizations can either over- or under-estimate trabeculae throughout the thresholding process if the thresholded region is sufficiently large. Choice of thresholding

technique has been the focus of several studies in the past and has shown that different thresholding techniques affect different parameters to different degrees, as much as ~ 25%⁴⁴, and parameters can also be affected by the observer⁴⁶. Thresholding remains one of the most important considerations for the microarchitectural characterisation of trabecular bone.

Several previous studies have compared morphometrics from μ CT to other imaging modalities, including cone beam CT (CBCT), multislice CT and MRI. In most cases, comparison was simply by correlation of parameters for each of the two imaging modalities^{47–50}. Van Dessel et al. and Klintström et al. showed many parameters were significantly different from those measured by μ CT, perhaps most importantly those parameters often used in bone disease research (e.g., BV/TV, TbTh, FD). Several studies reported morphometric values from either CBCT or MRI that were significantly different to corresponding μ CT values, and consequently, both studies relied on the correlation of parameters between imaging techniques rather than comparison of absolute values^{47,48,51}.

In this study, BA/TA values are consistent with literature values for the bovine femur and tibia, as are mean TbTh and TbSp values^{33,36,37}. A previous study reported highly variable FD for different regions of bovine rib⁵² but FD observed within this study is consistent with those seen in both human and non-human studies previously^{8,11,52}. Treating morphometrics measured by the μ CT as a ground truth, it is apparent that FCG can precisely determine BA/TA and FD in most cases. Calculated FD has been seen to vary with different box size⁵³ and different signal to noise ratios⁵⁴, which may explain the different values seen between different pixel sizes (e.g., FCG versus downsampled FCG and μ CT).

It can be seen that FCG can accurately measure FD, for an equivalent pixel size. Additionally, while FCG (particularly downsampled) accurately measures TbSp compared to μ CT, it does not have the same accuracy measuring TbTh. This is potentially due to the circle fitting method imageJ uses, and potentially convoluted further by the relatively large step in z (1 mm).

While clearly the ability of FCG to image all microarchitectural characteristics accurately is affected by the decreased resolution available in the z axis, the potential for future exploitation of the geometry is profound. Simultaneous absorption and diffraction imaging using FCG has tremendous implications for the future of fracture prediction once both structural and physicochemical information is available.

A potential limitation of this study is the resolution of the μ CT images; if a smaller voxel size could be achieved (i.e. by use of a smaller sample), morphometric results may be more precise. Though all efforts have been made to allow easy comparison with other imaging techniques used for morphometric analysis, some limitations are present. The presence of out-of-plane objects within the FCG image preclude this analysis from including traditional parameters commonly used in the clinical sphere such as TMD and ρ BMD. Microarchitectural characteristics measured are also limited to features within the 2D images investigated, and in the future a better average may be achieved by use of the 3D volume. This study has made every effort to follow guidelines accepted by the community for measurement of microarchitectural characteristics⁴⁵.

Conclusions

This study has shown that tomosynthesis using FCG presents an accurate measurement of microarchitectural morphometric parameters. FCG tomography allows for collection of both absorption and diffraction simultaneously⁵⁵. The combination of absorption and diffraction FCG would provide the simultaneous interrogation of both morphometric and physicochemical bone quality via a single scan using the same hollow beam. Thus, given that such quality measures are known to be significantly altered with bone disease^{9,56}, a combined absorption/diffraction probe promises significant clinical diagnostic utility.

Data availability

Data supporting this study are included within the article and/or supplementary materials.

Received: 19 June 2023; Accepted: 30 November 2023

Published online: 06 December 2023

References

1. Kanis, J. A. *et al.* SCOPE 2021: A new scorecard for osteoporosis in Europe. *Arch. Osteoporos.* <https://doi.org/10.1007/s11657-020-00871-9/Published> (2021).
2. Svedbom, A. *et al.* Osteoporosis in the European Union: A compendium of country-specific reports. *Arch. Osteoporos.* **8**, 33 (2013).
3. Farlay, D. & Boivin, G. Bone mineral quality. In *Osteoporosis* (ed. Dionysiotis, Y.). 3–32. <https://doi.org/10.5772/29091> (InTech, 2012).
4. Kanis, J. A., Johnell, O., Oden, A., Johansson, H. & McCloskey, E. FRAX and the assessment of fracture probability in men and women from the UK. *Osteoporos. Int.* **19**, 385–397 (2008).
5. Yerramshetty, J. S. & Akkus, O. Changes in cortical bone mineral and microstructure with aging and osteoporosis. In *Skeletal Aging and Osteoporosis* (ed Silva, M. J.) 105–131 <https://doi.org/10.1007/8415> (2013).
6. Teo, J. C. M., Si-Hoe, K. M., Keh, J. E. L. & Teoh, S. H. Relationship between CT intensity, micro-architecture and mechanical properties of porcine vertebral cancellous bone. *Clin. Biomech.* **21**, 235–244 (2006).
7. Seeman, E. & Delmas, P. D. Bone quality—The material and structural basis of bone strength and fragility. *N. Engl. J. Med.* **354**, 2250–2261 (2006).
8. Arnold, E. L., Clement, J. G., Rogers, K. D., Garcia-Castro, F. & Greenwood, C. The use of μ CT and fractal dimension for fracture prediction in osteoporotic individuals. *J. Mech. Behav. Biomed. Mater.* **103**, 103585 (2020).
9. Greenwood, C. *et al.* Towards new material biomarkers for fracture risk. *Bone* **93**, 55–63 (2016).
10. Greenwood, C. *et al.* Age-related changes in femoral head trabecular microarchitecture. *Aging Dis.* **9**, 976 (2018).
11. Alberich-Bayarri, A. *et al.* Assessment of 2D and 3D fractal dimension measurements of trabecular bone from high-spatial resolution magnetic resonance images at 3-T. *Med. Phys.* **37**, 4930–4937 (2010).
12. Audran, M., Chappard, D., Legrand, E., Libouban, H. & Basle, M. F. Bone microarchitecture and bone fragility in men: DXA and histomorphometry in humans and in the orchidectomized rat model. *Calcif. Tissue Int.* **69**, 214–217 (2001).

13. Cheung, W. *et al.* Best performance parameters of HR-pQCT to predict fragility fracture: Systematic review and meta-analysis. *J. Bone Miner. Res.* **36**, 2381–2398 (2021).
14. Klose-Jensen, R. *et al.* High-resolution peripheral quantitative computed tomography for bone evaluation in inflammatory rheumatic disease. *Front. Med. (Lausanne)* **7**, 45 (2020).
15. Brüllmann, D. & Schulze, R. K. W. Spatial resolution in CBCT machines for dental/maxillofacial applications—What do we know today?. *Dentomaxillofac. Radiol.* <https://doi.org/10.1259/dmfr.20140204> (2015).
16. Krug, R. *et al.* Feasibility of in vivo structural analysis of high-resolution magnetic resonance images of the proximal femur. *Osteoporos. Int.* **16**, 1307–1314 (2005).
17. Chesnut, C. H. *et al.* Effects of salmon calcitonin on trabecular microarchitecture as determined by magnetic resonance imaging: Results from the QUEST study. *J. Bone Miner. Res.* **20**, 1548–1561 (2005).
18. Engelke, K., Chaudry, O. & Bartenschlager, S. Opportunistic screening techniques for analysis of CT scans. *Curr. Osteoporos. Rep.* **21**, 65–76. <https://doi.org/10.1007/s11914-022-00764-5> (2023).
19. Chen, H. Y. *et al.* Application of tomosynthesis for vertebral compression fracture diagnosis and bone healing assessment in fracture liaison services. *Front. Med. (Lausanne)* **9**, 31 (2022).
20. Evans, J. P. O. *et al.* X-ray absorption tomography employing a conical shell beam. *Opt. Exp.* **24**, 29048 (2016).
21. Elarnaut, F. *et al.* Sporadic absorption tomography using a conical shell X-ray beam. *Opt. Exp.* **25**, 33029 (2017).
22. Dicken, A. J. *et al.* X-ray diffraction from bone employing annular and semi-annular beams. *Phys. Med. Biol.* **60**, 5803–5812 (2015).
23. Evans, P., Rogers, K., Dicken, A., Godber, S. & Prokopiou, D. X-ray diffraction tomography employing an annular beam. *Opt. Exp.* **22**, 11930 (2014).
24. Prokopiou, D., Rogers, K., Evans, P., Godber, S. & Dicken, A. Discrimination of liquids by a focal construct X-ray diffraction geometry. *Appl. Radiat. Isotopes* **77**, 160–165 (2013).
25. Li, F., Liu, Z. & Sun, T. Annular beam high-intensity X-ray diffraction based on an ellipsoidal single-bounce monocrystalline. *J. Appl. Crystallogr.* **49**, 627–631 (2016).
26. Greenwood, C. *et al.* Towards medical diagnostics using hyperspectral X-ray scatter. In *Advanced Materials TechConnect Briefs 2017*. Vol. 1. 32–35 (2017).
27. Evans, J. P. O. *et al.* Sparse interleaved sampling for high resolution focal construct geometry X-ray tomography. *Opt. Exp.* **31**, 15301 (2023).
28. MATLAB version: 9.10.0 (R2021a), Natick, Massachusetts: The MathWorks Inc. (2021).
29. Dobbins, J. T. Tomosynthesis imaging: At a translational crossroads. *Med. Phys.* **36**, 1956–1967. <https://doi.org/10.1118/1.3120285> (2009).
30. Kita, A. *et al.* Usefulness of top-hat transform processing in whole body bone scintigraphy. *Nihon Hoshasen Gijutsu Gakkai Zasshi* **69**, 41–48 (2013).
31. Guerroudj, M. A. & Ameer, Z. A new approach for the detection of mammary calcifications by using the white Top-Hat transform and thresholding of Otsu. *Optik (Stuttg)* **127**, 1251–1259 (2016).
32. Sawagashira, T. *et al.* An automatic detection method for carotid artery calcifications using top-hat filter on dental panoramic radiographs. *IEICE Trans. Inf. Syst.* **E96-D**, 1878–1881 (2013).
33. da Silva, A. M. H., Alves, J. M., da Silva, O. L. & da Silva Junior, N. F. Two and three-dimensional morphometric analysis of trabecular bone using X-ray microtomography (μ CT). *Rev. Bras. Engenharia Biomed.* **30**, 93–101 (2014).
34. Liu, C. *et al.* Relationships of ultrasonic backscatter with bone densities and microstructure in bovine cancellous bone. *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* **65**, 2311–2321 (2018).
35. Nagaraja, S., Lin, A. S. P. & Guldborg, R. E. Age-related changes in trabecular bone microdamage initiation. *Bone* **40**, 973–980 (2007).
36. Deligianni, D. D. & Apostolopoulos, K. N. Characterization of dense bovine cancellous bone tissue microstructure by ultrasonic backscattering using weak scattering models. *J. Acoust. Soc. Am.* **122**, 1180–1190 (2007).
37. Syahrom, A., Abdal Kadir, M. R., Abdullah, J. & Öchsner, A. Mechanical and microarchitectural analyses of cancellous bone through experiment and computer simulation. *Med. Biol. Eng. Comput.* **49**, 1393–1403 (2011).
38. Bradley, D. & Roth, G. Adaptive thresholding using the integral image. *J. Graph. Tools* **12**, 13–21 (2007).
39. Doube, M. *et al.* BoneJ: Free and extensible bone image analysis in ImageJ. *Bone* **47**, 1076–1079 (2010).
40. Domander, R., Felder, A. A. & Doube, M. BoneJ2—Refactoring established research software. *Wellcome Open Res.* **6**, 37 (2021).
41. Schindelin, J. *et al.* Fiji: An open-source platform for biological-image analysis. *Nat. Methods* **9**, 676–682. <https://doi.org/10.1038/nmeth.2019> (2012).
42. Hildebrand, T. & Rueggsegger, P. A new method for the model-independent assessment of thickness in three-dimensional images. *J. Microsc.* **185**, 67–75 (1997).
43. BenSaïda, A. *Shapiro–Wilk and Shapiro–Francia Normality Tests*. <https://uk.mathworks.com/matlabcentral/fileexchange/13964-shapiro-wilk-and-shapiro-francia-normality-tests> (2014).
44. Tassani, S., Korfiatis, V. & Matsopoulos, G. K. Influence of segmentation on micro-CT images of trabecular bone. *J. Microsc.* **256**, 75–81 (2014).
45. Bouxsein, M. L. *et al.* Guidelines for assessment of bone microstructure in rodents using micro-computed tomography. *J. Bone Miner. Res.* **25**, 1468–1486 (2010).
46. Rovaris, K. *et al.* Segmentation methods for micro CT images: A comparative study using human bone samples. *Braz. Dent. J.* **29**, 150–153 (2018).
47. Panmekiate, S., Ngonphloy, N., Charoenkarn, T., Faruangsang, T. & Pauwels, R. Comparison of mandibular bone microarchitecture between micro-CT and CBCT images. *Dentomaxillofac. Radiol.* **44**, 13 (2015).
48. Krug, R. *et al.* Assessment of trabecular bone structure comparing magnetic resonance imaging at 3 Tesla with high-resolution peripheral quantitative computed tomography ex vivo and in vivo. *Osteoporos. Int.* **19**, 653–661 (2008).
49. Van Dessel, J. *et al.* Accuracy and reliability of different cone beam computed tomography (CBCT) devices for structural analysis of alveolar bone in comparison with multislice CT and micro-CT. *Eur. J. Oral Implantol.* **10**, 95–105 (2017).
50. Klintström, E., Klintström, B., Spånges, A., Sandborg, M. & Woisetschlager, M. Trabecular bone microstructure analysis on data from a novel twin robotic X-ray device. *Acta Radiol.* <https://doi.org/10.1177/02841851221134973> (2022).
51. Tsai, M. T., He, R. T., Huang, H. L., Tu, M. G. & Hsu, J. T. Effect of scanning resolution on the prediction of trabecular bone microarchitectures using dental cone beam computed tomography. *Diagnostics* **10**, 88 (2020).
52. Perrotti, V. *et al.* Correlation between bone density and fractal dimension: A pilot study. *Nonlinear Phenomena Complex Syst.* **23**, 130–132 (2020).
53. Parkinson, I. & Fazzalari, N. Fractal analysis of trabecular bone: A standardised methodology. *Image Anal. Stereol.* **19**, 45–49 (2000).
54. Chung, H. W., Chu, C. C., Underweiser, M. & Wehrli, F. W. On the fractal nature of trabecular structure. *Med. Phys.* **21**, 1535–1540 (1994).
55. Shevchuk, A. *et al.* Combined X-ray diffraction and absorption tomography using a conical shell beam. *Opt. Exp.* **27**, 21092 (2019).
56. Dicken, A. J. *et al.* Classification of fracture and non-fracture groups by analysis of coherent X-ray scatter. *Sci. Rep.* **6**, 1–7 (2016).

Acknowledgements

The work was partly supported by an Engineering & Physical Sciences Research Council grant (EP/T034238/1). This work was also partly supported by a Royal Society Wolfson Fellowship (RSWF/R1/180012).

Author contributions

K.D.R., E.L.A., J.P.O.E., D.D., and F.E. conceived the idea and conducted the experiment. E.L.A. analysed the data and wrote the manuscript. K.D.R., J.P.O.E., D.D., F.E. and C.G. revised the manuscript.

Competing interests

The authors declare no competing interests.

Additional information

Supplementary Information The online version contains supplementary material available at <https://doi.org/10.1038/s41598-023-48851-6>.

Correspondence and requests for materials should be addressed to E.L.A.

Reprints and permissions information is available at www.nature.com/reprints.

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2023



DSPACE

<https://dspace.org/>

Conical shell X-ray beam tomosynthesis and micro-computed tomography for microarchitectural characterisation

Arnold, Emily; Elarnaut, Farid; Downes, David; Evans, J. Paul O.; Greenwood, Charlene ...

2023-12-06

Springer

<https://doi.org/10.1038/s41598-023-48851-6>

Arnold EL, Elarnaut F, Downes D, et al., (2023) Conical shell X-ray beam tomosynthesis and micro-computed tomography for microarchitectural characterisation. Scientific Reports, Volume 13, 2023, Article Number 21480